

CERTIFICATE OF CALIBRATION

THIS IS TO CERTIFY THAT THE SHIMPO INSTRUMENT IDENTIFIED BELOW HAS BEEN TESTED AT OUR FACTORY IN ACCORDANCE WITH APPROVED SHIMPO PROCEDURES. THE STANDARDS USED IN THE CALIBRATION OF THIS UNIT ARE PERIODICALLY INSPECTED AND ARE TRACEABLE TO THE JAPAN NATIONAL RESEARCH LABORATORY OF METROLOGY.

MODEL : FGV-20XY SERIAL NO. _____
DATA CERTIFIED: 02/15/2008 CAPACITY: 10kg × 0.01kg
ACCURACY: $\pm 0.2\%$ OF FULL SCALE $\pm 1/2$ LEAST SIGNIFICANT COUNT

EQUIPMENT MODEL: **DIGITAL MULTI METER 756202**
DATA CALIBRATED: **04/13/2006**
CALIBRATION DUE: **04/12/2008**
OFFICIAL INSPECTION: **SENSHIN ELECTRIC CO.,LTD.**

WEIGHTS USED:

**A-34 4000g, A-35 4000g, B-5 200g, B-6 400g, B-7 500g, B-61 400g, B-71 500g,
B-72 500g, B-73 500g, B-74 500g, B-75 500g, B-76 500g, B-77 500g**

DATA CALIBRATED: **04/02/2007**

CALIBRATION DUE: **04/01/2009**

OFFICIAL INSPECTION: **EACH WEIGHT IS CALIBRATED BY
ELECTRONIC BALANCE PM-6100.AND MP-300**

CALIBRATED by

寺 柿