


Factory Certificate of Calibration

Product Name	LOGIC BASIC	Name of Inspection Standard	Ig2110
Model No.		Unit	inch
Serial No		Scale Interval	0.0001 inch
Certificate No.		Measuring Range	1 inch
		Reference Point	0 inch
		End Point	1 inch


N.I.S.T. No. 821/268795-03

Inspection Item Name	Result	Permissible Value	Judgment
Indication Error	0.0000611 inch	0.0002 inch	GO
Hysteresis	-----	0.0001 inch	N/A
Repeatability	0.000002 inch	0.0001 inch	GO
Max. Measuring Force	-----	-----	N/A

Inspection Item Name	Judgment
Inspection of Function and Appearance	GO



Indication Error is the sum of accuracy and quantizing error.



Repeatability is taken at three positions, with five readings at each position.